
Temperature calibrator TP 17650M TP Basic // Dry block // Room temperature...650 °C // RT...1202 °F



TP 17650M - Highlights

- Version of our temperature calibrator TP 17650 specially optimised for the marine market
- Very easy operation with 4-button control and integrated reference temperature sensor
- Low weight and stable handle for easy transport
- Optional accessory: Transport bag or transport case with or without trolley
- Qualified for SIKA Gold Service

TP Basic

The temperature calibrators of the TP Basic series are characterised by their efficiency and portability. The series consists of dry block calibrators covering a wide temperature range and are used on site, e.g. in marine applications.

The easy operation, the integrated internal reference temperature sensor and the dry block calibration function ensure an extremely easy calibration process.

SIKA temperature calibrators

Temperature calibrators are used for the verification of the functionality and calibration of temperature measuring devices and temperature sensors. As the sole German manufacturer of these devices, we develop and produce our "Made in Germany" temperature calibrators with a special focus on long-term reliability and utmost accuracy in combination with easy operation. We can rely on more than 40 years of experience in doing this: SIKA's first dry block temperature calibrator was launched all the way back in 1980.

Every SIKA temperature calibrator is meticulously tested for accuracy and stability. This is attested by our standard calibration certificate, which we issue with every temperature calibrator, or by means of an optional DAkkS calibration certificate [German accreditation body]. This is to guarantee that you receive a perfect product which can be traced back to national and international temperature measurement standards.



Features

Easy operation

- The TP 17650M can be operated with only four buttons:
 Two arrow buttons for setting the target temperature,
 one button for confirmation and one return button
- Thus, temperatures can be set as easily as, for example, in the air conditioning system in your car
- Any operational errors can be nearly excluded. You do not need any specifically trained staff or time-consuming briefings



SIKA Gold Service

SIKA Gold Service provides a comprehensive service package for the regular recalibration of your temperature calibrator. You will benefit from exclusive savings and discounts as well as special promotions reserved to SIKA Gold Service members.

- You will save 33% in the recalibration of your temperature calibrator
- You will receive a 10% discount on any repairs that may become necessary
- You will receive preferential invitations to product presentations, symposia, practice days and exclusive training offers





Technical data

TD 45/5014				
TP 17650M				
Temperature range	Room temperature450 °C Room temperature1202 °F			
Dimension of the calibration insert	Ø 28 x 150 mm (calibration insert easily exchangeable)			
Dry block				
Display accuracy	±1 °C	±1.8 °F		
Temperature stability	±0.1 °C ±0.18 °F			
Resolution of the temperature display	1 °C 1 °F			
Reference temperature sensor	internal, fixed installation			
Dimensions				
→ Width→ Height→ Depth	150 mm 330 + 70 mm (Handle) 270 mm			
Weight	Approx. 7.5 kg			
Power supply → Standard → Optional	230240 VAC, 50 / 60 Hz 100115 VAC, 50 / 60 Hz 100240 VAC, 50 / 60 Hz			
Power consumption	Approx. 1000 W			
Display				
Display	2-line, 4-digit digital display red / green, unit °C / °F			
Approvals				
	ON THE COMM			

Article numbers

To order a complete calibrator, you need two article numbers:

- 1. Calibrator
- 2. Calibration insert

In addition, depending on your individual calibration requirements, you can order additional calibration inserts, necessary certificates and other accessories.

1. Calibrator					
Temperature range		Function	Calibration insert [mm]	Power supply	Article number
Room temperature650 °C	RT1202 °F	Dry block	Ø 28 x 150	230 V	EP17650M281500
Room temperature650 °C	RT1202 °F	Dry block	Ø 28 x 150	115 V	EP17650M281502
Room temperature650 °C	RT1202 °F	Dry block	Ø 28 x 150	110240 V	EP17650M281503

2. Calibration insert				
Bore holes [mm]	Function	Calibration insert [mm]	Material	Article number
1x Ø 3.5, 1x Ø 6.5, 1x Ø 13.5	Dry block	Ø 28 x 150	Brass	EZ15028B03MS17
1x Ø 6.5	Dry block	Ø 28 x 150	Brass	EZ15028065MS00
2x Ø 3.5	Dry block	Ø 28 x 150	Brass	EZ15028B02MS09
1x Ø 3.5, 1x Ø 4.5	Dry block	Ø 28 x 150	Brass	EZ15028F02MS80
1x Ø 3.5, 1x Ø 6.5	Dry block	Ø 28 x 150	Brass	EZ15028H02MS01
1x Ø 3.5, 1x Ø 8.5	Dry block	Ø 28 x 150	Brass	EZ15028B02MS67
1x Ø 3.5, 1x Ø 6.5, 1x Ø 8.5, 1x Ø 10.5	Dry block	Ø 28 x 150	Brass	EZ15028C04MS15
Without bore holes	Dry block	Ø 28 x 150	Brass	EZ15028000MS00
Calibration insert incl. 1 bore hole of choice	Dry block	Ø 28 x 150	Brass	Please indicate bore
Each additional bore hole	Dry block	Ø 28 x 150	Brass	holes in the order

3. Calibration certificate - Select your calibration certificates as needed Each calibrator is already delivered with a standard calibration certificate (3 test points).	Article number
SIKA works calibration certificate (similar to standard calibration certificate + marking on the calibrator)	EKTPWP1FKT
DAkkS calibration certificate (3 test points + measurement uncertainty determination)	EKTPDAKKS1FKT
Each additional test point DAkkS calibration certificate	EKTPDAKKSZUSP
SIKA Gold Service works calibration certificate	EKTPGOLDWP
SIKA Gold Service DAkkS	EKTPGOLDDAKKS

4. Accessories	Article number		
Transport case without trolley	EZTPKOFFER020		
Transport case with trolley	EZTPKOFFER020TG		
Transport bag	XE2193		



Overview of SIKA temperature calibrators

Temperature range (RT=Room temperature)	Function	Accuracy		Features	Block dimensions [Ø mm x depth mm]	Туре
-55 °C 200 °C -67 °F 392 °F	Dry block	±0.4 °C	±0.72 °F		28 x 150	TP 17200
	Dry block	±0.2 °C	±0.36 °F	PC interface	28 x 150	TP 17200S
	Dry block	±0.2 °C	±0.36 °F	Touch screen PC interface External reference sensor Integrated measuring instrument	28 x 150	TP 37200E.2
	Dry block	±1 °C	±1.80 °F		28 x 150	TP 17165M
	Dry block	±0.4 °C	±0.72 °F		28 x 150	TP 17165
	Dry block	±0.2 °C	±0.36 °F	PC interface	28 x 150	TP 17165S
	Dry block	±0.2 °C	±0.36 °F	Touch screen PC interface External reference sensor Integrated measuring instrument	28 x 150	TP 37165E.2
-35 °C 165 °C -31 °F 329 °F	Dry block	±0.4 °C	±0.72 °F		60 x 150	TP 17166
5 <u>52</u> , .	Dry block	±0.2 °C	±0.36 °F	PC interface	60 x 150	TP 17166S
	Calibration bath	±0.1 °C	±0.18 °F	PC interface	60 x 170	TP M165S
	Dry block Air Shield Insert Calibration bath Infrared Surface	±0.3 °C ±0.099 °C ±0.1 °C ±0.5 °C ±1 °C	±0.54 °F ±0.1782 °F ±0.18 °F ±0.9 °F ±1.88 °F	Touch screen PC interface External reference sensor Integrated measuring instrument	60 x 170	TP 3M165E.2
-10 °C 100 °C 14 °F 212 °F	Dry block	±0.05 °C	±0.09 °F	PC interface	7 x 6.5 x 150	TP 17Zero
RT 200 °C RT 392 °F	Dry block	±1 °C	±1.80 °F		18 x 150	TP 18200E
	Calibration bath	±0.2 °C	±0.36 °F	PC interface	60 x 170	TP M255S
RT 255 °C RT 491 °F	Dry block Calibration bath Infrared Surface	±0.3 °C ±0.2 °C ±0.5 °C ±1°C	±0.54 °F ±0.36 °F ±0.9 °F ±1.8 °F	Touch screen PC interface External reference sensor Integrated measuring instrument	60 x 170	TP 3M255E.2
	Dry block	±0.6 °C	±1.08 °F		60 x 150	TP 17450
	Dry block	±0.3 °C	±0.54 °F	PC interface	60 x 150	TP 17450S
RT 450 °C RT 842 °F	Dry block Air Shield Insert Infrared Surface	±0.3 °C ±0.2 °C ±0.5 °C ±1 °C	±0.54 °F ±0.36 °F ±0.9 °F ±1.8 °F	Touchscreen PC interface External reference sensor Integrated measuring instrument	60 x 150	TP 37450E.2
DT /F0.00	Dry block	±1 °C	±1.80 °F		28 x 150	TP 17650M
RT 650 °C RT 1202 °F	Dry block	±0.8 °C	±1.44 °F		28 x 150	TP 17650
	Dry block	±0.4 °C	±0.72 °F	PC interface	28 x 150	TP 17650S
RT 700 °C RT 1292 °F	Air Shield Insert	±0.53 °C	±0.954 °F	Touchscreen PC interface External reference sensor Integrated measuring instrument	29 x 150	TP 37700E.2
RT 850 °C RT 1562 °F	Dry block	±1 °C	±1.80 °F		18 x 100	TP 18850E
400 °C 1300 °C 752 °F 2372 °F	Dry block	±2 °C	±3.6 °F	PC interface	28 x 200	TP 281300E

Subject to technical modifications and errors

